

Quick Fringe Test Report



PO Box 2313
Iowa City, IA 52244-2313
www.opticalmechanics.com
jcmulherin@opticalmechanics.com

Comments: Serial Number 12-127-062508
F=62.5 inches

Output Units: Waves

RMS 0.034
Strehl Ratio 0.96

Scale -0.50
Wavelength 550.0 nm
Description OMI 12.5 inch F/5
Date Tested June 25, 2008
Operator JCM
Comments Serial Number 12-127-062508

