

Quick Fringe Test Report



PO Box 2313
Iowa City, IA 52244-2313
www.opticalmechanics.com
jcmulherin@opticalmechanics.com

Comments: Serial# 15-276-092308a
F=67.1 inches

Output Units: Waves

RMS 0.032
Strehl Ratio 0.96

Description OMI 15 inch F/4.5
Date Tested Sept 23, 2008
Operator James Mulherin
Comments Serial# 15-276-092308a

